

# REVOLON

Integrated scan controller with scan generator and image acquisition for transmission electron microscopes (TEM)

## User manual

Document version: 2.0

Date of issue: 2026-03-04

Document language: EN

Translation of German original

**Use** Anybody working with/on the device has to have read and understood the relevant parts of this document.

**Access** The staff working with the device has to have constant access to this document to prevent handling errors and guarantee trouble-free operation.

**Reproduction** All information in this document is protected by copyright. This document must not be copied, distributed or changed in any way by anyone including business units or departments of point electronic GmbH without prior written approval by point electronic GmbH. Reproduction and use of this document are confined to internal purposes of the operator only.

**Up-to-datedness** Every endeavor has been made to make sure that the information contained in this document is complete and correct at the printing date. This document describes all units and functions known of at the current point of time.

**Brand, company, or product names** The brand, company, or product names used in this document are the brand, company, or product names of the respective manufacturers and/or owners.

**Contact us** point electronic GmbH  
Erich-Neuß-Weg 15  
06120 Halle (Saale)  
Germany  
Phone: +49 345 1201190  
Email: [info@pointelectronic.de](mailto:info@pointelectronic.de)  
Web: <https://www.pointelectronic.de>

**Product information** Product: REVOLON  
Description: Integrated scan controller with scan generator and image acquisition for transmission electron microscopes (TEM)  
Name: PE-REVOLON-TEM-SCAN  
Manufacturer: point electronic GmbH

# Table of contents

<b>1</b>	<b>Introduction</b> .....	<b>5</b>
	User groups of this document .....	6
	Structure of this document .....	7
	Representations in this document .....	8
	Identification of the warning notices .....	9
	Topics relevant to service .....	10
<b>2</b>	<b>Safety regulations</b> .....	<b>11</b>
	Introduction .....	12
	Basic hazards .....	13
	Staff and qualifications .....	14
	Responsibility of the operating company .....	15
	Working in a safety-conscious manner .....	16
	Modifications and alterations .....	17
	Maintenance work .....	18
	Cleaning .....	20
	Environmental protection .....	21
<b>3</b>	<b>System overview</b> .....	<b>22</b>
	Intended use .....	23
	Improper use .....	24
	Scope of delivery .....	25
	System environment requirements .....	26
	Front panel .....	27
	Back panel .....	28
	Touch display .....	31
	Device labeling .....	34
<b>4</b>	<b>Pinouts of the inputs and outputs</b> .....	<b>36</b>
	Analog I/O .....	37
	100 MHz digital .....	39
	Sync .....	40
	IN B1-B4 .....	41
	Scan in .....	42
	Scan out .....	43
<b>5</b>	<b>Installation and configuration</b> .....	<b>44</b>
	Configuring the inputs and outputs .....	45
	Mounting in a 19-inch rack .....	48
	Connecting to the power supply .....	49
<b>6</b>	<b>Maintenance</b> .....	<b>50</b>
	Replacing the fuses .....	51
<b>7</b>	<b>Disposal</b> .....	<b>53</b>
	Recycling and taking back of used equipment .....	54
<b>8</b>	<b>Specifications</b> .....	<b>55</b>

Operating conditions ..... 56  
Device specifications ..... 57



# 1 Introduction

---

## Chapter overview

**Purpose** This chapter contains information to simplify working with this document.

**Contents** This chapter contains the following information:

- › User groups of this document . . . . . 6
- › Structure of this document . . . . . 7
- › Representations in this document . . . . . 8
- › Identification of the warning notices . . . . . 9
- › Topics relevant to service . . . . . 10

## User groups of this document

**Operators** "Operators" are persons with right of disposal over the device using this right for purchase or other purposes.

**Technical staff** "Technical staff" are those persons that the operators entrust with tasks related to use and operation. Technical staff are trained by the operators to carry out their assigned tasks and are informed on the potential hazards that may arise due to improper handling.

Technical staff need to undergo a training by point electronic GmbH or a partner company of point electronic GmbH covering the following topics:

- Handling of the connected TEM
- Areas of application and operation of the device and the associated software
- Handling the connected peripherals
- Basic maintenance tasks and troubleshooting

**Service technicians** "Service technicians" are staff members whose specialist training, knowledge and experience as well as familiarity with the relevant safety regulations mean they are able to assess the tasks they have been entrusted with and anticipate potential hazards.

Service technicians may be:

- Employees of point electronic GmbH or
- Employees of partner companies of point electronic GmbH

## Structure of this document

- Composition** This document is composed of chapters which are organized by technical aspects.
- Numbering** The chapters are numbered with Arabic numerals. Chapters may be composed of sections. Sections are numbered as second numbering level (e. g. 3.1). Sections are used to structure large chapters into sub-chapters.
- All pages of this document are numbered consecutively.
- Overviews** Each chapter and section contains an overview detailing contents and page numbers. This allows for direct access to a topic and independent use of single parts of this document.
- Related information** Related information within this document is marked by the note "Continuation next page ..." resp. "... continuation:". Please pay attention to the completeness of the information when copying parts of this document.
- Cross references** The content of this document is structured by topics. If further information on one topic may be found elsewhere in the user guide, the relevant chapter and page are pointed out.

## Representations in this document

**Illustrations** Illustrations used in this document do not always contain all details or special cases. They only represent the relevant information.

**Icons** The following icons are used to mark certain information visually:

Icon	Meaning
	Notice For example:  The selected parameter will not be inserted in the parameter order.
	Reference to another part of this guidebook For example:  see "Cross references" on page 7
	Use of a tool For example:  Screwdriver TX 10

**Notes** Important notes are marked as follows:

### NOTICE

#### Mind the notices in this document!

Notes explain relations that even for expert users might not be evident at first glance.

The neglect of a note is no direct security risk. However, it can lead to disturbances in the operating procedure.

## Identification of the warning notices

**Purpose** This document includes warning notices that may lead to serious consequences if disregarded. Warning notices are not only listed in the “Safety regulations” chapter, but especially in places where hazards for people, equipment and operation may arise.

**Identification of warnings** There are three classes of warnings. These classes are indicated by specific signal words and colors. They include the following:

Signal word	Meaning
 <b>DANGER</b>	Warning notice, which if disregarded will probably or very likely result in death or serious injury.
 <b>WARNING</b>	Warning notice, which if disregarded may result in serious injury, permanent damage to health or serious property loss.
 <b>CAUTION</b>	Warning notice, which if disregarded may result in injury or property loss including financial losses due to operational impairment.
 <b>ATTENTION</b>	Warning notice, which if disregarded may result in property loss including financial losses due to operational impairment.

## Topics relevant to service

**Service contact** For detailed information and questions about topics such as installation and configuration, please contact our service department at:

[service@pointelectronic.de](mailto:service@pointelectronic.de)

**Further documents** Additional documents with descriptions and instructions for installation, configuration, and additional software tools are available in protected public areas.

If you require access to further documents, please contact our service department at:

[service@pointelectronic.de](mailto:service@pointelectronic.de)



## 2 Safety regulations

---

### Chapter overview

**Purpose** This chapter contains safety information relating to the protection of persons as well as safe and fault-free operation. All user groups of the hardware and software must be aware of and follow these safety provisions.

**Contents** This chapter contains the following information:

› Introduction . . . . .	12
› Basic hazards . . . . .	13
› Staff and qualifications . . . . .	14
› Responsibility of the operating company . . . . .	15
› Working in a safety-conscious manner . . . . .	16
› Modifications and alterations . . . . .	17
› Maintenance work . . . . .	18
› Cleaning . . . . .	20
› Environmental protection . . . . .	21

## Introduction

- Reliable and safe operation** The reliable and safe operation of the device depends on carefully implementing the operating, setup and maintenance tasks.
- Observance of the safety instructions** Always observe and follow the safety instructions and accompanying code of conduct when handling the device and the software. Always point out the instructions and code of conduct to staff working with the hardware and software.
- Generally applicable safety regulations (such as accident prevention and environmental protection regulations, etc.) must also be observed.
- Consequences arising from failure to observe the safety instructions** Failure to observe the safety instructions can result in death or serious injury of personnel and damage or destruction of the hardware components.

## Basic hazards

**Definition** The device meets the current state of the art as well as the applicable safety regulations. The device has been tested at the manufacturer's site and is delivered in a state safe for operation.

Basic hazards are residual risks that may arise despite proper and safety-conscious use of the device.

### **DANGER**

#### **Risk of death by electrocution!**

Body contact with live parts may lead to death or serious injuries.

- Turn off the mains voltage before any works on the device.
- Secure the device against unintentional restart.
- Works on the device may only be executed by service technicians.
- Never operate the device when the mains cable is damaged.

### **WARNING**

#### **Risk of injury due to improper use!**

Improper use of the device may lead to injury.

- Always protect the device from extreme heat (excessive sun exposure, close proximity to open fire or heating equipment) during operation and storage.
- Avoid hard impacts that might damage parts of the device.

## Staff and qualifications

**Permissions** Works on/with the device may only be performed by technical staff and/or service technicians (🔑 see "User groups of this document" on page 6).

Note the legally prescribed minimum age for personnel.

Only qualified electricians or technical staff under guidance and supervision of a qualified electrician may perform works on electrical components of the device in accordance with the electrotechnical regulations.

Staff undergoing training or instructions or persons taking part in general vocational training programs may only operate the device under the continuous supervision of technical staff.

## Responsibility of the operating company

**Condition of the device and/or overall system** The operating company shall ensure that all safety and the protective devices are fully functional at all times. This equipment should therefore be checked periodically for functionality and completeness.

This applies to the device as single component as well as the overall system into which the device is integrated as a subcomponent.

**Internal measures to avoid hazards** The operating company must be aware of the applicable industrial safety regulations and use a risk assessment to determine any additional hazards that may arise from the special working conditions at the device's site of operation. Internal instructions for avoiding the identified hazards should then be drawn up.

Throughout the entire operating period of the device, the operating company must examine and determine whether the supplied operating instructions comply with the current status of applicable regulations.

**Staff** Regarding the staff authorized or trained by the operating company, the operating company carries the following responsibilities:

- The necessary training and instruction of staff must be guaranteed.
- The powers and responsibilities of staff for installation, operation, maintenance and service must be clearly defined and documented.
- This document must be kept in the immediate surroundings of the device and must be readily accessible to staff.

## Working in a safety-conscious manner

**Accident prevention and environmental protection** In addition to the information in this document, also note the generally applicable statutory and other binding regulations relating to accident prevention and environmental protection.

This may include, for example:

- dealing with hazardous substances
- wearing the necessary and prescribed personal protective clothing
- observance of and compliance with all national and regional industrial safety regulations
- observance of and compliance with all internal working, operating and safety regulations

**Content of this document** Prior to operation, all staff members assigned to work on the device must have read and understood the relevant parts of this document, especially the chapter “Safety regulations”.

## Modifications and alterations

**No unauthorized modifications** Any unauthorized modifications to the device exclude any liability by the manufacturer for resulting damage and consequences.

Do not perform any modifications, additions and/or alterations to the device without written authorization from point electronic GmbH.

**Spare parts and accessories** Spare parts and accessories must comply with the technical requirements specified by point electronic GmbH and its suppliers. Whenever original parts are used, compliance is given.

**No modifications of the software** No modifications to the software must be made or commissioned to third parties. The software may not be deleted, decrypted or decompiled in full or in part.

If changes to the software are necessary, please contact point electronic GmbH.

## Maintenance work

**Implementation** High reliability as well as low maintenance costs of the device may be achieved through careful implementation of the maintenance schedule and regular monitoring over the entire operational lifetime.

**Compliance with regulations** When performing maintenance work, it is essential to observe the following:

- applicable accident prevention regulations,
- applicable environmental protection regulations and
- safety instructions for maintenance work.

Only perform maintenance work if:

- you are authorized to do so and
- the device is turned off and secured against unintentional restart.

**Before getting started** Before you begin maintenance work, please make sure that:

- the device is de-energized and disconnected from the power supply and
- the power supply may not be re-established unintentionally as long as the maintenance work lasts.

**Works during operation** In any case, the operating company or the personnel employed by it must check based on the specific local conditions that the specified work may be performed during operation without endangering anybody.

Continuation next page ...

**Electrical equipment** Regularly check the electrical equipment of the device. Immediately rectify defects (such as loose cable connections and/or faulty cables).

Damaged or faulty cables may only be replaced by cables that match the requirements stipulated by point electronic GmbH and/or its suppliers!

Check that the surfaces of all electrical equipment parts are dry and free of oil, grease, deposits and corrosion.

Only use voltage-insulated tools!

Do not place any tools or working materials on the conductive surfaces of components.

**Device fuses** Faulty fuses may only be replaced by fuses that match the requirements (specifications and ratings) stipulated by point electronic GmbH!

Faulty fuses may not be repaired or bypassed, but must be replaced by fuses of the same type.

## Cleaning

- Suitable cleaning products** Only use cleaning products approved by point electronic GmbH.
- All cleaning products used by the customer should be checked with point electronic GmbH or the respective supplier to ensure they are compatible with the materials and colors used.
- Unsuitable cleaning agents** Do not clean the device with:
- scratchy, aggressive cleaning agents containing solvents, gasoline, or alcohol,
  - pressured air, high-pressure cleaner or other kinds of cleaning machine.
- Cleaning equipment** Use lint-free cleaning cloths/wipes for cleaning the device.
- To remove stubborn dirt, use lint-free cleaning cloths/wipes moistened with clear water.
- After cleaning** After cleaning, make sure that:
- cables, ports and fittings are free of oil and/or cleaning agents and
  - cables, wires, connectors and electrical components are dry.

## Environmental protection

**Recyclable materials** When performing maintenance work, ensure that reusable materials are recycled.

**Disposal** Electrical and electronic waste may constitute a hazard to health and environment when disposed off improperly. According to WEEE Directive 2012/19/EU (Waste Electrical and Electronic Equipment Directive), electrical and electronic waste may not be disposed off as general domestic waste but must be handed in at specific collection points or sent back to the manufacturer.

Check with your local environmental protection agency for the prescribed disposal options for commercially used electronic waste.

Only hire waste disposal companies that are approved by the national and regional authorities.



## 3 System overview

---

### Chapter overview

**Purpose** This chapter describes the device. It contains information on the use, scope of delivery, equipment, structure and function.

**Contents** This chapter contains the following information:

› Intended use . . . . .	23
› Improper use . . . . .	24
› Scope of delivery. . . . .	25
› System environment requirements . . . . .	26
› Front panel . . . . .	27
› Back panel . . . . .	28
› Touch display . . . . .	31
› Device labeling . . . . .	34

## Intended use

**Purpose** The device and software are exclusively intended for the acquisition and management of digital image data.

The device is designed for installation and operation in a 19-inch rack.

**Operation** The device may only be operated in a technically perfect condition and as intended, in compliance with the user manual and under consideration of safety and potential hazards. Malfunctions, in particular those that may impair safety, should be rectified immediately.

If the device shows damages or defects that compromise operational safety, this must immediately be reported to the operations center, and the device must not be put into operation.

Operation and maintenance of the device may only be performed by technical staff and service technicians (🔑 see "User groups of this document" on page 6). The valid safety and accident prevention regulations must be observed.

**Compliance with regulations** All operating, maintenance, setup and service measures prescribed by the manufacturer must be complied with.

**Additional information** In addition to this document, the generally applicable, legal and other binding regulations and legislation and environmental protection instructions must be adhered to.

## Improper use

Any use not authorized by the manufacturer is not permissible and may lead to injury or property damage.

point electronic GmbH does not accept liability for damage arising from improper use of the device.

In particular, improper use includes:

- operation in explosive environments
- operation in an environment that does not meet the stipulated requirements
  -  see “Operating conditions” on page 56
- operation with opened device housing
- operation with visible damage to the device and/or peripherals (e.g. connected cables, etc.)
- modifications and/or additions that impair the performance
- exchanging components with unauthorized components

## Scope of delivery

The following basic equipment is part the scope of delivery:

Component	Item number
REVOLON TEM scan controller (device)	depends on the type of device delivered
Mains cable	
USB cable (Type-A to Type-B)	
Set of connecting cables for scan and video	
USB stick with: <ul style="list-style-type: none"><li>– Documentation,</li><li>– USB device driver and</li><li>– Sample codes for Python, C and C#</li></ul>	

### NOTICE

The complete scope of delivery with the particular item numbers depends on the respective order.


## System environment requirements

**Minimum requirements** The following table contains information on the minimum requirements of the system environment:

Component	Requirement
Computer	IBM-compatible from Core i3 Recommendation: Core i5 <hr/> Resolution at least 1280×1024 pixel, true color Recommended resolution: 1920×1200 pixel <hr/> At least one free USB 2.0 or USB 3.0 slot Optional: one free LAN slot (RJ-45) for TCP/IP connection
Power connection	At least one socket: <ul style="list-style-type: none"><li>– 115 ... 230 V AC</li><li>– 50 ... 60 Hz</li><li>– Single-phase</li><li>– Same grounding as the microscope</li></ul>
TEM connection	<ul style="list-style-type: none"><li>– 1 × external scan interface (daisy-chain configuration supported)</li><li>– At least one video signal output</li></ul>
Mounting	Space of 2 U in a 19 inch rack

## Front panel

**Structure** The following figure shows the front of the device with its components:




**Components** The following table contains information on the components on the front:

No.	Component
1	19" mounting brackets for installation in a rack
2	INPUTS - 5 MHz ANALOG <ul style="list-style-type: none"> <li>– 4, 8 or 16 image signal input ports</li> <li>– max. <math>\pm 5</math> V</li> <li>– <math>\pm 0.5</math> V (with gain = 1x)</li> <li>– gain: 1 ... 1800x</li> <li>– input offset: <math>\pm 1.25</math> V</li> <li>– type: SMB, socket (female)</li> </ul>
3	Touch display <ul style="list-style-type: none"> <li>– displays status and system information for the device</li> <li>– enables various configurations for the device</li> </ul> ⓘ see "Touch display" on page 31

## Back panel

**Structure** The following figure shows the back of the device with its components:



**Components** The following table contains information on the components of the back:

No.	Component
1	<p>SCAN OUT</p> <ul style="list-style-type: none"> <li>– differential scan output</li> <li>– switched to REVOLON scan if a REVOLON scan is running</li> <li>– type: 9 pin, D-Sub, socket (female)</li> </ul> <p> pinout: page 43</p> <p> configuration: page 45</p>
2	<p>ANALOG I/O</p> <ul style="list-style-type: none"> <li>– port for analog inputs and outputs</li> <li>– type: 25 pin, D-Sub, plug (male)</li> </ul> <p><b>Analog outputs</b></p> <ul style="list-style-type: none"> <li>– amplitude: max. <math>\pm 12</math> V</li> <li>– output offset: <math>\pm 2</math> V</li> </ul> <p><b>Analog inputs</b></p> <ul style="list-style-type: none"> <li>– max. <math>\pm 5</math> V</li> <li>– <math>\pm 0.5</math> V (with gain = 0 dB)</li> <li>– gain: <math>-22 \dots +26</math> dB</li> <li>– input offset: <math>\pm 1.25</math> V</li> </ul> <p> pinout: page 37</p>
3	<p>SCAN IN</p> <ul style="list-style-type: none"> <li>– differential scan input</li> <li>– connected to SCAN OUT if no scan is running</li> <li>– type: 9 pin, D-Sub, plug (male)</li> </ul> <p> pinout: page 42</p>
4	<p>DVI</p> <ul style="list-style-type: none"> <li>– port</li> <li>– type: DVI-I (dual link), socket (female)</li> </ul>

Continuation next page ...

No.	Component
5	<p>LAN</p> <ul style="list-style-type: none"> <li>– port for TCP/IP control interface</li> <li>– transmission of commands and data</li> <li>– type: RJ-45, socket (female)</li> </ul>
6	<p>USB</p> <ul style="list-style-type: none"> <li>– port for USB control interface</li> <li>– transmission of commands and data</li> <li>– type: USB type B, socket (female)</li> </ul>
7	<p>PREAMP INTERFACE</p> <ul style="list-style-type: none"> <li>– signal input 20 Bit</li> <li>– control interface (RS-485) for additional amplifier</li> <li>– type: RJ-45, socket (female)</li> </ul>
8	<p>INPUTS - 100 MHz ANALOG</p> <ul style="list-style-type: none"> <li>– 4 image signal input ports</li> <li>– <math>\pm 5</math> V max.</li> <li>– <math>\pm 0.5</math> V (with gain = 0 dB)</li> <li>– gain: -22 ... +26 dB</li> <li>– <math>\pm 1.25</math> V input offset</li> <li>– type: BNC, socket (female)</li> </ul>
9	<p>IN B1-4</p> <ul style="list-style-type: none"> <li>– port for additional analog inputs (fast)</li> <li>– 4 image signal inputs</li> <li>– multiplexed B inputs to "100 MHz ANALOG"</li> <li>– max. <math>\pm 5</math> V</li> <li>– <math>\pm 0.5</math> V (with gain = 0 dB)</li> <li>– gain: -22 ... +26 dB</li> <li>– input offset: <math>\pm 1.25</math> V</li> <li>– type: 9 pin, D-Sub, plug (male)</li> </ul> <p> pinout: page 41</p>
10	<p>SYNC</p> <ul style="list-style-type: none"> <li>– trigger input and output signals</li> <li>– input: 3.3/5 V TTL</li> <li>– output: 5 V TTL</li> <li>– type: 15 pin, D-Sub, socket (female)</li> </ul> <p> pinout: page 40</p>
11	<p>100 MHz DIGITAL</p> <ul style="list-style-type: none"> <li>– 12 counting inputs</li> <li>– 3.3/5 V TTL counter</li> <li>– type: 25 pin, D-Sub, socket (female)</li> </ul> <p> pinout: page 39</p>
12	<p>Mains voltage connector</p> <ul style="list-style-type: none"> <li>– 115 ... 230 V AC, 25 W, 50 ... 60 Hz</li> <li>– type: C14, male</li> </ul>
13	<p>Grounding connector</p>

Continuation next page ...

No.	Component
14	Fuses – 2 pieces – replaceable – type: T4A
15	On/off switch

## Touch display

**Description** The touch display shows status and device information on four switchable pages and allows various configurations for the device.


**Status** The Status page displays status information about the current scan.

The following figure shows the touch display with the Status page:



**Option** The Options page shows an overview of the installed hardware options and the size of the pixel map.

The following figure shows the touch display with the Option page:



Continuation next page ...

**Info** The Info page displays information about the installed firmware and current device parameters.

The following figure shows the touch display with the Info page:

REVOLON by point electronic	
Firmware	1.56, 224/82/114/106
Vcc(P)Int (1.0V)	0.99 (0.98)V
Vcc(P)Aux (1.8V)	1.8 (1.79)V
VccDDR (1.5V)	1.49V
Temp Die/Board	53.1/43.5 Degrees
+/- 15V	15.21/-15.12V
+/-5V	5.05/-5.06V
+3.3V (digital)	3.32V
+3.3V (analog)	3.33V
+5V (digital)	5.09V
2.5V	2.52V
Scan X/Y	Not Inverted, Inverted
Mics UART	0

Buttons: Status, Option, **Info**, TCP, DVI

**TCP** The TCP page shows an overview of the current network configuration and allows enabling/disabling of TCP and DHCP.

The following figure shows the touch display with the TCP page:


REVOLON by point electronic	
TCP	Enabled
DHCP	Disabled
IP	192.168.0.11
MAC	00-0A-35-01-7C-09
Mask	255.255.255.0
Gate	192.168.1.1
SUP	7701
DSCP	7702
DIIP	7703
DIDP	7704
DVDP	7705
State	Wait for Clients

Buttons: Status, Option, Info, **TCP**, DVI

Continuation next page ...

**DVI** On the page DVI, the DVI output of the device can be configured. This is useful if a monitor is to be connected directly to the device.

The following figure shows the touch display with the DVI page:




The following options are available for configuring the DVI output:

Option	Description
DVI Monitor	Enabling/disabling the DVI output.
Resolution	Setting the screen resolution.
Source	Setting the input signal source to be displayed on the screen.

## Device labeling

**Markings on the device** The following figure shows the markings on the back of the device:




The following table contains information on the markings on the back of the device:

No.	Component
1	Grounding marking
2	Identification plate
3	Information on power supply, power consumption and fuses

Continuation next page ...

**Identification plate** The following figure shows the identification plate with its components:



The following table contains information on the components of the identification plate:

No.	Component
1	Manufacturer identification
2	Disposal symbol in accordance with the WEEE directive
3	CE marking
4	Manufacturer address
5	Serial number
6	Device designation

## 4 Pinouts of the inputs and outputs

---

### Chapter overview

**Purpose** This chapter describes the pinouts of the particular inputs and outputs.



**Contents** This chapter contains the following information:

› Analog I/O.....	37
› 100 MHz digital.....	39
› Sync.....	40
› IN B1-B4 .....	41
› Scan in.....	42
› Scan out.....	43

## Analog I/O

**Type** 25 pin, D-Sub, plug (male)

**Numbering** The following table contains information on the numbering of the connections for the respective shape:

Shape	Numbering (view from outside)
Plug	
Socket	

**Assignment** The following table contains information on the assignment of the connection numbers:

No.	Assignment
1	Scan Y out
2	Scan X out
3	Ext. scan out
4	Ext. blank out
5	Image signal A1+
6	Image signal A1-
7	Image signal A2+
8	Image signal A2-
9	Image signal A3+
10	Image signal A3-
11	Image signal A4+
12	Image signal A4-
13	Not connected
14	Scan Y in
15	Scan X in
16	Ext. scan in
17	Ext. blank in



Continuation next page ...

<b>No.</b>	<b>Assignment</b>
18...24	GND
25	Not connected
Housing	Shield

## 100 MHz digital

**Type** 25 pin, D-Sub, socket (female)

**Numbering** The following table contains information on the numbering of the connections for the respective shape:

Shape	Numbering (view from outside)
Plug	
Socket	



**Assignment** The following table contains information on the assignment of the connection numbers:

No.	Assignment
1	X 1
2	X 2
3	X 3
4	X 4
5	X 5
6	X 6
7	X 7
8	X 8
9	X 9
10	X 10
11	X 11
12	X 12
13	Not connected
14...25	GND
Housing	Shield

## Sync

**Type** 15 pin, D-Sub, socket (female)

**Numbering** The following table contains information on the numbering of the connections for the respective shape:

Shape	Numbering (view from outside)
Plug	
Socket	



**Assignment** The following table contains information on the assignment of the connection numbers:

No.	Assignment	Type
1	PixelClock	output: 5 V TTL
2	LineClock	output: 5 V TTL
3	FrameClock	output: 5 V TTL
4	HoldClock	output: 5 V TTL
5	Ext. scan	output: 5 V TTL
6	Ext. blank	output: 5 V TTL
7	BeamBlanker	
8	Lock-In ref out	
9	GND	
10	GND	
11	Pixel sync	input: 3.3/5 V TTL
12	Line sync	input: 3.3/5 V TTL
13	Frame sync	input: 3.3/5 V TTL
14	Hold sync	input: 3.3/5 V TTL
15	Spare I/O	
Housing	Shield	

## IN B1-B4

**Type** 9 pin, D-Sub, plug (male)

**Numbering** The following table contains information on the numbering of the connections for the respective shape:

Shape	Numbering (view from outside)
Plug	
Socket	



**Assignment** The following table contains information on the assignment of the connection numbers:

No.	Assignment
1	Image signal B1+
2	Image signal B2+
3	Image signal B3+
4	Image signal B4+
5	GND
6	Image signal B1-
7	Image signal B2-
8	Image signal B3-
9	Image signal B4-
Housing	Shield

## Scan in

**Type** 9 pin, D-Sub, plug (male)

**Numbering** The following table contains information on the numbering of the connections for the respective shape:

Shape	Numbering (view from outside)
Plug	
Socket	



**Assignment** The following table contains information on the assignment of the connection numbers:

No.	Assignment
1	Scan Y-
2	Not connected
3	Not connected
4	Not connected
5	Scan X-
6	Scan Y+
7	Not connected
8	Not connected
9	Scan X+
Housing	Shield

## Scan out

**Type** 9 pin, D-Sub, socket (female)

**Numbering** The following table contains information on the numbering of the connections for the respective shape:

Shape	Numbering (view from outside)
Plug	
Socket	

**Assignment** The following table contains information on the assignment of the connection numbers:

No.	Assignment
1	Scan Y-
2	Not connected
3	Not connected
4	Not connected
5	Scan X-
6	Scan Y+
7	Not connected
8	Not connected
9	Scan X+
Housing	Shield



## 5 Installation and configuration

---

### Chapter overview

**Purpose** This chapter contains descriptions and instructions on installation and configuration of the device.

**Contents** This chapter contains the following information:

- › Configuring the inputs and outputs . . . . . 45
- › Mounting in a 19-inch rack . . . . . 48
- › Connecting to the power supply . . . . . 49

## Configuring the inputs and outputs

**Description** On the carrier board of the device, the inputs and outputs of the device are adjusted to the requirements of the connected TEM.

### **⚠ DANGER**

#### **Risk of death by electrocution!**

Body contact with live parts may lead to death or serious injuries as well as the destruction or damaging of components.

- Prior to the configuration of inputs and outputs, please make sure that the device is de-energized and secured against unintentional starting (restart).


**Before getting started** Complete the following steps before configuring the inputs and outputs of the device:

1. Switch off the device.
2. Remove the mains cable.
3. Remove all connected cables from the device.

**Steps** Complete the following steps to configure the inputs and outputs of the device:

1. Remove the 4 screws (TORX TX 10) on the back of the device.

 screwdriver TX 10




2. Push back the upper casing cover.

Continuation next page ...

... Continuation: Configuring the inputs and outputs

3. Configure the inputs and outputs.
  -  see "Connector strips and assignment" on page 46
  -  see "Configurations for K13 and K14" on page 47
4. Configure the differential scan outputs.
  -  see "Configuration of the differential scan outputs" on page 47
5. Close the casing cover.
6. Attach the 4 screws (TORX TX 10) on the back of the REVOLON.
  -  screwdriver TX 10

**Connector strips and assignment** The following figure shows the distribution of the connector strips and their standard assignment:



The following table contains information on function and assignment of the connector strips:


Connector strip	Function	Assignment
K18	Terminating resistor for signal inputs A1 ... A4 and B1 ... B4	Jumper closed = terminated with 50 Ω
K11, K12	Type of scan output	<ul style="list-style-type: none"> <li>– Scan (standard) = fixed configured signal level</li> <li>– Mag = signal level adjustable with DAC</li> </ul>
K16, K17	Type of signal input A4	<ul style="list-style-type: none"> <li>– A4 (standard) = analog signal</li> <li>– Vid: configuration for composite video (PAL)</li> </ul>
K13, K14	Configuration of outputs "ext. Scan" and "ext. Blank"	<ul style="list-style-type: none"> <li>– K13 = signal for ext. Scan</li> <li>– K14 = signal for ext. Blank</li> </ul>

Continuation next page ...

**Configurations for K13 and K14**


The digital outputs "ext. Scan" and "ext. Blank" are potential-free relay outputs destined for switching the TEM to external beam control. They must be adapted to the requirements of the TEM using the K13 and K14 connector strips.

The following figure shows the possible configurations of the "ext. Scan" and "ext. Blank" outputs on the connector strips K13 and K14:



**Configuration of the differential scan outputs**

The voltage range of the differential scan outputs must be adapted to the requirements of the TEM as follows:



Voltage range	Configuration
Jumper set: ±0,65 ... ±2,2 V	
Jumper not set: ±2,2 ... ±7,5 V	

## Mounting in a 19-inch rack

**Installation site** Ensure that the installation site adheres to the stipulated operating conditions.

 see "Operating conditions" on page 56

**19-inch rack** Ensure that the 19-inch rack is equipped with horizontal mounting rails for each built-in device.

**Mounting the device** Complete the following steps to mount the device in a 19-inch rack:

### WARNING

#### Risk of injury when lifting, holding and carrying the device!

The weight of the device can lead to injuries and/or damage to health when lifting, holding and carrying it.


- Always assemble/disassemble the device in pairs.
- Wear the prescribed protective work clothing (e.g. safety shoes).

### NOTICE

#### Always ensure accessibility!

When mounting the device, please ensure that on-off switch and all connectors are accessible without obstructions.

1. Place the device in a 19-inch rack.
  -  Mind the installation height of two rack units (U).
  -  **Ensure that the device rests on the horizontal mounting rails in the rack.**
2. Fix the device to the rack frame with four screws.



3. Tighten the four screws hand-tight.

## Connecting to the power supply

**Steps** Complete the following steps to connect the device to the power supply:


### NOTICE

#### Mind the device connectors!


When connecting the device, please refer to the description of the unit's backside.

 see "Back panel" on page 28

1. Connect the grounding to the device.



2. Connect the AC power cable to the device.



 **Use the AC power cable included in the scope of delivery or an AC power cable that meets the same requirements!**

3. Connect the AC power cable to a power supply.

 **The power supply must comply with the specified requirements!**

 see "Device specifications" on page 57



## 6 Maintenance

---

### Chapter overview

**Purpose** This chapter contains instructions on the maintenance of the device.

**Contents** This chapter contains the following information:  
› Replacing the fuses . . . . . 51

## Replacing the fuses

**Position of the fuses** The fuses are located at the back of the device.

 see "Back panel" on page 28

### **DANGER**

#### **Risk of death by electrocution!**

Body contact with live parts may lead to death or serious injuries as well as the destruction or damaging of components.

- Prior to the replacement of the fuses, please make sure that the device is de-energized and secured against unintentional starting (restart).


**Before getting started** Complete the following steps before replacing the fuses of the device:

1. Switch off the device.
2. Remove the mains cable.
3. Remove all connected cables from the device.

Continuation next page ...

**Steps** Complete the following steps to replace the fuses of the device:

1. Pinch the locking tongues of the fuse holder and pull the fuse holder from the slot.



2. Replace the fuse or the fuses.  
**⚠ Only use type T4A fuses!**
3. Push the fuse holder into the slot until the locking tongues snap in.



# 7 Disposal

---

## Chapter overview

**Purpose** This chapter contains information on the disposal of the device.

**Contents** This chapter contains the following information:  
› Recycling and taking back of used equipment . . . . . 54

## Recycling and taking back of used equipment

### Disposal in accordance with regulations

Electrical and electronic devices may pose a risk to health and the environment if disposed of incorrectly. They cannot therefore be disposed of as domestic waste according to WEEE Directive 2012/19/EU (Waste Electrical and Electronic Equipment Directive). Instead they must be taken to designated collecting points or returned to the manufacturer.

Electrical and electronic devices must undergo specified recycling processes (e.g. with respect to circuit boards) which enable safe, environmentally compatible re-use or separate disposal of different device elements.

The return of used devices is regulated differently in different places. Find out from your local council about the return conditions for commercially used electronic devices.

The device contains no toxic substances requiring separate identification for disposal such as mercury (Hg), cadmium (Cd), lead (Pb) or chrome 6 (e.g. in circuit boards).

### Labeling

The following symbol indicates the legal duty to dispose of electronic devices as stipulated:





## 8 Specifications

---

### Chapter overview

**Purpose** This chapter lists the specifications of the device.

**Contents** This chapter contains the following information:

- › Operating conditions ..... 56
- › Device specifications ..... 57

## Operating conditions

The following table contains information on the operating conditions of the device:

<b>Site of operation</b>	<ul style="list-style-type: none"><li>– indoor</li><li>– permanently installed device (mounted in a 19-inch rack)</li><li>– controlled electromagnetic environment</li></ul>
<b>Ambient air</b>	<ul style="list-style-type: none"><li>– average air pressure (approx. 1 bar)</li><li>– dry</li><li>– free of dust</li></ul>
<b>Ambient temperature</b>	5 ... 40°C
<b>Humidity</b>	max. 80% (at 31°C, non-condensing)
<b>Altitude</b>	max. 2,000 m

## Device specifications

**Electrical safety** The following table contains information on electrical safety:

<b>Protection class</b>	I
<b>Protection type</b>	IP20
<b>Overvoltage category</b>	II
<b>Contamination level</b>	1

**Power supply** The following table contains information on the specifications of the power supply:

<b>Supply voltage</b>	115 ... 230 V AC
<b>Frequency</b>	50 ... 60 Hz
<b>Power consumption</b>	max. 25 W
<b>Fluctuation of the supply voltage</b>	max. $\pm 10\%$

**Fuses** The following table contains information on the fuses:


<b>Number</b>	2 pc. (replaceable)
<b>Type</b>	T4A
<b>Rating</b>	4 A slow blowing

**Dimensions and weight** The following table contains information on the dimensions and weights:

<b>Height</b>	88.1 mm (2 U)
<b>Width</b>	482.6 mm (incl. 19-inch mounting brackets)
<b>Depth</b>	290 mm
<b>Weight</b>	approx. 4.7 kg

Continuation next page ...

**Dimensional drawing** The following figure shows the dimensions of the device:



Continuation next page ...

**Device housing** The following table contains information on the components of the device housing:

<b>Cover plate and base plate</b>	Material: Aluminum Surface: powder-coated RAL 7035 (light gray)
<b>Front panel and rear panel</b>	Material: Aluminum Surface: powder-coated RAL 7035 (light gray)
<b>Frame and profiles</b>	Material: Aluminum Surface: powder-coated RAL 7001 (silver gray)
<b>19-inch mounting brackets</b>	Material: Die-cast aluminum Surface: powder-coated RAL 7001 (silver gray), contact points bare